

TABLA DE BIOMETRÍA - ACRIVUE

ESPECIFICACIÓN		
FÓRMULA	BIOMETRÍA POR ULTRASONIDO	BIOMETRÍA ÓPTICA
SRK II	118.0	119.1
SRK T	117.8	118.8
Hoffer Q	pACD-4.80	pACD-5.39
Haigis	$\alpha 0-0.586$	$\alpha 0-1.177$
	$\alpha 1-0.400$	$\alpha 1-0.400$
	$\alpha 2-0.100$	$\alpha 2-0.100$
Hollady I	sf-1.02	sf-1.61