

TABLA DE BIOMETRÍA – AQUAFOLD

ESPECIFICACIÓN		
FÓRMULA	BIOMETRÍA POR ULTRASONIDO	BIOMETRÍA ÓPTICA
SRK II	118.2	118.65
SRK T	118.0	118.4
Hoffer Q	pACD-4.91	pACD-5.14
Haigis	$\alpha 0-0.673$	$\alpha 0-0.912$
	$\alpha 1-0.400$	$\alpha 1-0.400$
	$\alpha 2-0.100$	$\alpha 2-0.100$
Hollady I	sf-1.12	sf-1.35